

## ADDITIONAL INFORMATION NO. 2

**Public Contract name:** High Capacity Roughing Pump System

**Contracting Authority:** Fyzikální ústav AV ČR, v.v.i.  
**Registered office:** Na Slovance 2, 182 21 Prague 8  
**Company identification No.** 68378271  
**Tax identification No.** CZ68378271

(hereinafter referred to as the "Contracting Authority")

In Prague on 30<sup>th</sup> October 2015

Above mentioned Contracting Authority pursuant to Sec. 49 of the Act n. 137/2006 Coll., on Public Contracts (hereinafter as "Act") is announcing Additional Information regarding the Tender Documentation of the above mentioned below-the-threshold Public Contract on Supplies awarded in open procedure pursuant to Sec. 27 of the Act.

### Question

**REQ-010001/A**

**The maximum size of the pump system including the frame shall be 1.8 m x 2 m x 2.3 m (high).**

**Can the length be instead of 2 m 2,1 m long so 100mm longer?**

Answer: The length has to be less than 2 m, as mentioned in the technical requirement REQ-010001/A

Mgr. Zuzana Šenoldová  
OTIDEA a.s.