

ADDITIONAL INFORMATION NO. 1

Public Contract name: Direct Write Laser Lithography System

Contracting Authority: Institute of Photonics and Electronic of the CAS, v. v. i.
Registered office: Chaberská 57, 182 51 Praha 8 - Kobylisy
Company identification No. 67985882
Tax identification No. CZ67985882

Person authorised to act on behalf of the Contracting Authority: Prof. Ing. Jiří Homola, CSc., DSc., Director

(hereafter referred to as the "**Contracting Authority**")

In Prague on 06/04/2016

Above mentioned Contracting Authority pursuant to Sec. 49 of the Act n. 137/2006 Coll., on Public Contracts (hereinafter as "Act") is announcing additional information regarding the Tender Documentation of the above mentioned below-the-threshold Public Contract on supplies awarded in the form of simplified below-the-threshold awarding procedure.

Request:

Your documents state that an offer will automatically be disqualified if one of the technical requirements of Annex No. 4 is not fulfilled. What if we fail to meet one of the requirements but the system would still be able to fully meet the requirements in some other way?

To be more specific, the light source at 405nm is not necessarily required to meet your requirements. The light source at 375nm can expose all optical photoresists in the range of 365-430nm. Our technology offers the possibility to combine various laser spot sizes with only one installed laser. Therefore, we could meet your requirements to expose all optical resist with various spot sizes, but would have to say "No" to REQ_02.

Please let me know in an official statement if an offer will automatically be rejected if one of the specification in annex 4 is not fulfilled without consideration of the technical implications.

Answer:

The Contracting Authority hereby confirms that he insists on the above mentioned requirement REQ_02 "Light source on 405 nm" in the original scope.

The Contracting Authority insists on the requirement mainly due to achievement of requested quality of resulting structures.

The Contracting Authority confirms regarding that Annex no. 4 of the Tender Documentation the following:

Should the offered value for the requirements REQ_03 - REQ_12 be stated by the Bidder outside from the value stated in the column B, the bid shall be excluded for the reason of failing to meet the tender conditions.

Should the Bidder state for the requirements REQ_01, REQ_02, REQ_13- REQ_17 that the offered solution does not provide this functionality, the bid shall be excluded for the reason of failing to meet the tender conditions.

It follows that in case of not fulfilling the requirement REQ_02 "Light source on 405 nm" the bid shall be automatically rejected.

No tender conditions have been amended by this additional information.

Yours sincerely,

Mgr. Zuzana Šenoldová
OTIDEA avz s.r.o.